Comparison and correlation of commercial SARS-CoV-2 real-time-PCR assays, Ireland, June 2020

Anne Carroll¹, Eleanor McNamara¹

1. Public Health Laboratory-Health Service Executive-Dublin, Ireland

Correspondence: Anne Carroll (anne.carroll@hse.ie)

Citation style for this article:

Carroll Anne, McNamara Eleanor. Comparison and correlation of commercial SARS-CoV-2 real-time-PCR assays, Ireland, June 2020. Euro Surveill. 2021;26(6):pii=2002079. <https://doi.org/10.2807/1560-7917.ES.2021.26.6.2002079>

Article submitted on 11 Dec 2020 / accepted on 11 Feb 2021 / published on 11 Feb 2021

We report the performance of a variety of commercially available SARS-CoV-2 PCR kits, used in several different sites across Ireland to determine if C_t values across platforms are comparable. We also investigate whether a C_t value, a surrogate for calculated viral loads in the absence of viral culture of >34 can be used to exclude SARS-CoV-2 infection and its complications. We found a variation in C_t values from different assays for the same calculated viral load; this should be taken into consideration for result interpretation.

The interpretation of severe acute respiratory syndrome coronavirus 2 (SARS-CoV-2) real-time (RT)-PCR tests presents multiple conundrums with respect to viral load to cause infection and be infectious, age, clinical phase (presymptomatic, symptomatic, asymptomatic, resolution, re-infection or persistent positivity), testing purposes (diagnostic or surveillance), trending of previous test results if available and use of test result (e.g. for infection prevention control or occupational health purposes). In recent months, various publications have suggested that the use of cycle threshold (C_t) values as surrogate for calculated viral load, may help in the management of patients [1-4].

In this study, we investigate if C_t values obtained by a variety of commercially available SARS-CoV-2 PCR kits, used in several different sites across Ireland, are comparable across platforms. We also explore whether a C_t value of >34 [3], in the absence of viral culture, can be used to exclude SARS-CoV-2 infection.

Ethical statement

No patient data or specimens were used in this study, therefore, ethical approval was not required.

PCR platform comparison

In April 2020, Quality Control for Molecular Diagnostics (QCMD) produced a SARS-CoV-2 external quality assessment (EQA) panel [5]. The panel contained eight

samples of which five were positive for SARS-CoV-2. Laboratories participating in the EQA were given the panel of samples without the respective information on positivity or negativity. Each laboratory processed the samples as if the provided material was viral transport media (VTM) from a SARS-CoV-2 inoculated swab, and tested the EQA panel according to their RT-PCR of choice and own laboratory procedures, then returned a 'Detected' or 'Not detected' result to QCMD. Following submission of results from all participants, in June 2020, QCMD provided a report to all participants, now detailing the digital (d)PCR $\log_{10}$ copies/mL of SARS-CoV-2 in the samples for reference purpose. The dPCR $\log_{10}$ copies/mL for the five SARS-CoV-2 positive samples were 4.3, 3.3, 4.3, 5.3, and 2.3 respectively.

In this study, we analyse the results of 16 participating clinical diagnostic laboratories across Ireland in more detail, using in particular the C_t values that they obtained with their RT-PCR assays for each of the five positive samples. For each laboratory, the C_t values for the five samples and the corresponding dPCR $\log_{10}$ copies/mL were employed to produce standard curves for each assay (or for each assay target). When several laboratories used a common assay, this allowed to assess the performance of the same assay across the platforms. Moreover, when laboratories used different assays, it was possible to compare outputs across assays. Six of 16 laboratories submitted data on more than one assay. In total nine assays with in total 15 gene targets were analysed (Table 1).

There were a number of different RNA extraction systems, either prior to the RT-PCR or incorporated within this procedure (no external extraction for GeneXpert (Cepheid)). Each of these systems used a different amount of sample (200 μ L to 750 μ L), with a varying portion of the total recovered RNA as subsequent PCR template. A standard curve was created in Microsoft Excel for each assay (one for each gene target). The standard

TABLE 1

Real-time PCR assays considered in the study, Ireland, June 2020 (n=9 assays)

Assay	Number of laboratories	Number of gene targets	Genes targeted	Total datasets generated
GeneXpert (Cepheid, Sunnyvale, California, United States)	6	2	E gene N gene	12
Logix Smart (Co-Diagnostics, Inc, Salt Lake City, Utah, United States)	2	1	RdRp gene	2
Cobas 4800 (Roche Diagnostics, Basel, Switzerland)	2	2	ORF1a/b E gene	4
RealStar (Altona Diagnostics GmbH, Hamburg, Germany)	1	1	E gene	1
genesig (Primerdesign, Southampton, Hants, United Kingdom)	4	1	ORF1a/b	4
RespiBio (SeroSep, Limerick, Ireland)	2	1	RdRp gene	2
VIASURE (CerTest Biotec, Zaragoza, Spain)	3	2	ORF1a/b N gene	3 (2 genes combined)
Abbott RealTime SARS-CoV-2 (Abbott Park, Illinois, United States)	1	2	RdRp gene N gene	1 (2 genes combined)
Allplex SARS-CoV-2 (Seegene, Seoul, South Korea)	1	3	RdRp gene N gene E gene	3

E: envelope; N: nucleocapsid; ORF: open reading frame; RdRp: RNA-dependent RNA polymerase; SARS-CoV-2: severe acute respiratory syndrome coronavirus 2.

TABLE 2Average and standard deviation of Ct values at calculated SARS-CoV-2 viral loads of $3\log_{10}$ to $8\log_{10}$ copies/mL, Ireland, June 2020 (n=8 assays)^a

Assay		Average C_t @ $3\log_{10}$ (SD)	Average C_t @ $4\log_{10}$ (SD)	Average C_t @ $5\log_{10}$ (SD)	Average C_t @ $6\log_{10}$ (SD)	Average C_t @ $7\log_{10}$ (SD)	Average C_t @ $8\log_{10}$ (SD)
GeneXpert	E	32.8 (1.32)	29 (0.57)	26 (0.63)	22 (0.72)	19 (0.79)	15 (0.89)
	N	35 (1.34)	31 (0.27)	27 (0.21)	23 (0.20)	20 (0.22)	16 (0.28)
Logix Smart	RdRp	37.5 (0.70)	33 (0.29)	30 (0.14)	26 (0.05)	23 (0.20)	19 (0.40)
Cobas 4800	ORF1a/b	33 (0.19)	29 (0.02)	26 (0.03)	23 (0.13)	20 (0.2)	17 (0.27)
	E gene	33.5 (0.7)	30 (0.05)	27 (0.03)	24 (0.10)	22 (0.15)	18 (0.25)
RealStar ^b	E	31 (NA)	28 (NA)	25 (NA)	22 (NA)	19 (NA)	16 (NA)
genesig	ORF1a/b	36 (1.63)	33 (0.43)	29 (0.72)	26 (0.97)	22 (1.31)	19 (1.59)
RespiBio	RdRp	37 (1.41)	33 (0.03)	30 (0.22)	26 (0.54)	23 (0.79)	19 (1.12)
VIASURE	ORF1a/b N (combined)	37 (3.78)	33 (0.96)	30 (0.87)	26 (0.99)	23 (1.21)	19 (1.6)
Allplex SARS-CoV-2 ^b	E	32 (NA)	29 (NA)	27 (NA)	24 (NA)	22 (NA)	19 (NA)
	N	34 (NA)	32 (NA)	29 (NA)	26 (NA)	23 (NA)	21 (NA)
	RdRp	36 (NA)	32 (NA)	28 (NA)	24 (NA)	19 (NA)	16 (NA)

Ct: cycle threshold; E: envelope; NA: not applicable; N: nucleocapsid; ORF: open reading frame; RdRp: RNA-dependent RNA polymerase; SARS-CoV-2: severe acute respiratory syndrome coronavirus 2; SD: standard deviation.

^a While a total of nine SARS-CoV-2 real-time PCR assays were used in the study-participating laboratories, the Abbott RealTime SARS-CoV-2 assay reports cycle number values, which are not equivalent to C_t values and thus are not directly comparable. This assay was therefore excluded from the analysis.

^b Absolute values for Altona and Seegene targets as only data from one laboratory.

TABLE 3

Average SARS-CoV-2 viral loads of log₁₀ copies/mL and standard deviation at selected C_t values, Ireland, June 2020 (n = 8 assays)^a

C _t	Average log ₁₀ (standard deviation) (copies/mL)											
	GeneXpert (Cepheid)		Logix Smart (Co-Diagnostics Inc)	Cobas 4800 (Roche)		RealStar (Altona Diagnostics) ^b	genesig (Primerdesign)	RespiBio (SeroSep)	VIASURE (CerTest Biotec)	Allplex SARS-CoV-2 (Seegene)		
	E	N	RdPd	E	ORF1a/b	E	ORF1a/b	RdPd	ORF1a/b, N (combined)	Eb	Nb	RdRpb
30	3.84 (0.54)	4.23 (0.27)	4.99 (0.19)	3.63 (0.07)	4.10 (0.11)	3.50 (NA)	4.77 (0.56)	5.02 (0.14)	4.95 (0.88)	3.77 (NA)	4.64 (NA)	4.48 (NA)
31	3.55 (0.53)	3.97 (0.28)	4.70 (0.24)	3.30 (0.09)	3.76 (0.13)	3.19 (NA)	4.48 (0.49)	4.74 (0.05)	4.67 (0.92)	3.37 (NA)	4.27 (NA)	4.24 (NA)
32	3.26 (0.52)	3.70 (0.30)	4.42 (0.29)	2.97 (0.12)	3.41 (0.16)	2.87 (NA)	4.19 (0.43)	4.46 (0.03)	4.38 (0.96)	2.98 (NA)	3.90 (NA)	4.01 (NA)
33	2.97 (0.50)	3.43 (0.32)	4.14 (0.34)	2.64 (0.14)	3.07 (0.19)	2.55 (NA)	3.91 (0.38)	4.17 (0.11)	4.10 (1.02)	2.58 (NA)	3.54 (NA)	3.77 (NA)
34	2.69 (0.50)	3.17 (0.34)	3.86 (0.39)	2.31 (0.17)	2.72 (0.22)	2.24 (NA)	3.62 (0.35)	3.89 (0.19)	3.82 (1.09)	2.19 (NA)	3.17 (NA)	3.53 (NA)
35	2.40 (0.49)	2.90 (0.36)	3.58 (0.44)	1.98 (0.19)	2.38 (0.24)	1.92 (NA)	3.33 (0.34)	3.61 (0.27)	3.53 (1.17)	1.79 (NA)	2.81 (NA)	3.30 (NA)
36	2.11 (0.49)	2.63 (0.38)	3.30 (0.49)	1.65 (0.22)	2.04 (0.27)	1.61 (NA)	3.05 (0.35)	3.32 (0.36)	3.25 (1.26)	1.40 (NA)	2.44 (NA)	3.06 (NA)
37	1.83 (0.54)	2.37 (0.27)	3.02 (0.19)	1.32 (0.07)	1.69 (0.11)	1.29 (NA)	2.76 (0.56)	3.04 (0.14)	2.97 (0.88)	1.00 (NA)	2.08 (NA)	2.83 (NA)

C_t: cycle threshold; E: envelope; NA: not applicable; N: nucleocapsid; ORF: open reading frame; RdRp: RNA-dependent RNA polymerase; SARS-CoV-2: severe acute respiratory syndrome coronavirus 2.

^a While a total of nine SARS-CoV-2 real-time PCR assays were used in the study-participating laboratories, the Abbott RealTime SARS-CoV-2 assay reports cycle number values, which are not equivalent to C_t values and thus are not directly comparable. This assay was therefore excluded from the analysis.

^b Absolute values for Altona and Seegene targets as only data from one laboratory.

curve was created by plotting the C_t value for each of the five samples against the dPCR log₁₀ copies/mL provided by QCMD. R² values of all the assays ranged from 0.9497 to 0.9997, with a mean of 0.9885. From each standard curve, estimated viral load (log₁₀ copies per mL) was extrapolated (using the equation of the individual standard curve) for each C_t value. The Abbott RealTime SARS-CoV-2 assay reports cycle number (CN) values, which are not equivalent to C_t values and thus are not directly comparable [6] and was therefore excluded from further analysis.

Mean C_t values for all assays and standard deviations were calculated across a range of log₁₀ copies per mL values (Table 2 and 3). Where only one laboratory tested an assay, absolute values were used for comparison. Correlation of C_t results between the same assays used across different sites was good for all assays (mean: 1.6; standard deviation: 0–5.1) (Figure, Table 2). Data from all assays correlated with the internationally recognised 3.3 cycle difference for every 1 log₁₀ copies/mL change in viral load. However, there was a wide variation in C_t values for different assays for the same viral loads, 6.5 cycle difference (31–37.5) at 3 log₁₀ down to 4 cycle difference (22–26) at 6 log₁₀. But the range difference in C_t values between assays was stable across all log values (Figure). These data demonstrate that reporting C_t values per se can be misleading and is non comparative between different assays, unless the

C_t value is correlated with the calculated viral load for the particular assay used and also reported.

Discussion

Inferred viral loads

Tom et al. 2020 [3] noted that the issue of high C_t values can be problematic for clinicians, especially when there are less than 100 copies of the virus present, as this could reflect presymptomatic, early infection, late infection, persistent positivity or nonviable virus. From our data even at a C_t as high as 37, six of eight assays had at least one gene target correlating to calculated viral load of ≥100 copies/mL, although our study did not include viral culture, nor infer cultureability. The C_t cut off of 34 described by La Scola et al. [7] at which they propose patients can be discharged from isolation may need to be considered cautiously, as our data show that a C_t value of 34 has a range of calculated viral loads from 2.19 log₁₀ to 3.89 log₁₀ (equivalent to approximately 150–7000 viral copies per mL (Table 3).

Our National coronavirus disease (COVID-19) guidelines [1,2] also describe the difficulty of interpreting positive 'high' (>30) C_t PCR results from asymptomatic individuals, they too suggest that a C_t of 34 equates to <100 copies/mL, however, our data indicate that the

FIGURE

Average C_t values at SARS-CoV-2 viral loads of $3\log_{10}$ to $8\log_{10}$ copies/mL, Ireland, June 2020 ($n = 8$ assays)^a

C_t : cycle threshold; E: envelope; N: nucleocapsid; ORF: open reading frame; R: RNA-dependent RNA polymerase; SARS-CoV-2: severe acute respiratory syndrome coronavirus 2.

^a While a total of nine SARS-CoV-2 real-time PCR assays were used in the study-participating laboratories, the Abbott RealTime SARS-CoV-2 assay reports cycle number values, which are not equivalent to C_t values and thus are not directly comparable. This assay was therefore excluded from the analysis.

^b When more than one laboratory used an assay (or targets in the assay), the average C_t s obtained across the laboratories are presented for each assay (or assay target).

Average C_t values at calculated viral loads of $3\log_{10}$ to $8\log_{10}$ copies/mL.

C_t value could be up to 38 for a calculated viral load of 100 copies/mL (data not shown).

A number of different RNA extraction systems were used by the participating laboratories, the effect this has on the results is unquantifiable. While it is a limitation of this study, our results support the view of Chik-Yan et al. [8] who state that differences in C_t values may be due to differences in specimen source or preparation or differences in cycling parameters and reagents, even though there is no significant difference in sensitivity. Indeed, for our data there was good correlation in C_t results using the same assay at different sites. Another limitation of the current study is that it investigated a small number of laboratories, with only one to six laboratories using the same assay (giving either absolute C_t values or averages based on small number of replicates). A larger study in the future would be useful to support these results.

Consensus on the correlation between C_t value and disease severity has not been reached, Sang Hyun Ra et al. [9] found that there was no significant difference in mean C_t values from symptomatic or asymptomatic cases, whereas Salvatore et al. [4] reported higher C_t values in asymptomatic individuals. Likewise Prubelli et al. [10] identified an increase in C_t values that correlated with a decrease in severe cases in Italy. While diagnostic test results play a role in identification and aid management of infected individuals, it is imperative to have a thorough understanding of the performance characteristics of individual PCR assays to aid the accurate interpretation of results [11].

Cycle threshold values in patient management

Using C_t values to influence patient management is complex and must be done with caution. Including the C_t value on positive results may be confusing and

misleading [12]. With no clearly defined infectious dose for SARS-CoV-2, viral culture not available routinely and C_t values differing by up to 6.5 cycles between platforms, one would question the value of the routine use of reporting C_t values for patient management. This is particularly challenging in a community testing setting, where frequently there is no accompanying patient clinical information to aid interpretation. However, this may be ameliorated in an acute hospital testing environment with access to patient clinical data. In this setting, analysis of C_t trends (either rising or decreasing) from repeat testing using the same assay may give more insight to an individual's disease progression or resolution.

To aid the clinical interpretative value of COVID-19 PCR results, we agree with Tom et al. that binary reporting ('Detected' or 'Not detected') could be enhanced by the additional reporting of C_t value ranges in 'high', 'medium', 'low' categories. However, our data infer that such ranges if used, should be based on calculated viral loads of each assay used, not on absolute C_t values. The viral loads for 'high', 'medium' and 'low' categories would need to be defined. This reporting would be more accurate and informative to aid clinical and public health decisions, particularly when considered in the context of individual clinical data.

Acknowledgements

We would like to thank all contributing laboratories; The Coombe Women & Infants University Hospital, University Hospital Waterford, Our Lady's Hospital Navan, PHL-HSE-Dublin, Cork University Hospital, University Hospital Limerick, Midland Regional Hospital Portlaoise, University Hospital Kerry, Children's Health Ireland (CHI) at Crumlin, UCD National Virus Reference Laboratory, Letterkenny University Hospital, University Hospital Galway (UHG), The Meath Foundation Tallaght University Hospital, Tallaght, Midland Regional Hospital at Tullamore, The Mater Misericordiae University Hospital and Rotunda Hospital.

Conflict of interest

None declared.

Authors' contributions

Anne Carroll: study design, data collection, data analysis and manuscript preparation. Eleanor McNamara: study design, critically revised manuscript.

References

1. Health Service Executive (HSE). HSE 2020, Proposal for the management of weak positive (high C_t value) PCR results in the setting of mass testing of asymptomatic individuals for SARS-CoV-2, V1.1 08.10.2020. Ireland: HSE; Dec 2020. Available from: <https://www.hpsc.ie/az/respiratory/coronavirus/novelcoronavirus/guidance/outbreakmanagementguidance/PCR%20weak%20results%20guidance.pdf>
2. Health Service Executive (HSE) Health Protection Surveillance Centre. (HPSC). Guidance on the management of weak positive (high C_t value) PCR results in the setting of testing individuals for SARS-CoV-2 V1.2 22.12.2020. Ireland: HSE HPSC. Available

- from: <https://www.hpsc.ie/az/respiratory/coronavirus/novelcoronavirus/guidance/outbreakmanagementguidance/PCR%20weak%20results%20guidance.pdf>
3. Tom MR, Mina MJ. To Interpret the SARS-CoV-2 Test, Consider the Cycle Threshold Value. *Clin Infect Dis.* 2020;71(16):2252-4. <https://doi.org/10.1093/cid/ciaa619> PMID: 32435816
 4. Salvatore PP, Dawson P, Wadhwa A, Rabold EM, Buono S, Dietrich EA, et al. Epidemiological Correlates of Polymerase Chain Reaction Cycle Threshold Values in the Detection of Severe Acute Respiratory Syndrome Coronavirus 2 (SARS-CoV-2). *Clin Infect Dis.* 2020;ciaa1469. <https://doi.org/10.1093/cid/ciaa1469> PMID: 32986120
 5. Quality Control for Molecular Diagnostics (QCMD). [Accessed 11 Feb 2021]. Available from: <http://www.qcmd.org/>
 6. Marais G, Naidoo M, Hsiao NY, Valley-Omar Z, Smuts H, Hardie D. The implementation of a rapid sample preparation method for the detection of SARS-CoV-2 in a diagnostic laboratory in South Africa. *PLoS One.* 2020;15(10):e0241029. <https://doi.org/10.1371/journal.pone.0241029> PMID: 33079951
 7. La Scola B, Le Bideau M, Andreani J, Hoang VT, Grimaldier C, Colson P, et al. Viral RNA load as determined by cell culture as a management tool for discharge of SARS-CoV-2 patients from infectious disease wards. *Eur J Clin Microbiol Infect Dis.* 2020;39(6):1059-61. <https://doi.org/10.1007/s10096-020-03913-9> PMID: 32342252
 8. Yip CC, Sridhar S, Cheng AK, Leung KH, Choi GK, Chen JH, et al. Evaluation of the commercially available LightMix® Modular E-gene kit using clinical and proficiency testing specimens for SARS-CoV-2 detection. *J Clin Virol.* 2020;129:104476. <https://doi.org/10.1016/j.jcv.2020.104476> PMID: 32516739
 9. Ra SH, Lim JS, Kim GU, Kim MJ, Jung J, Kim SH. Upper respiratory viral load in asymptomatic individuals and mildly symptomatic patients with SARS-CoV-2 infection. *Thorax.* 2021;76(1):61-3. <https://doi.org/10.1136/thoraxjnl-2020-215042> PMID: 32963115
 10. Piubelli C, Deiana M, Pomari E, Silva R, Bisoffi Z, Formenti F, et al. Overall decrease in SARS-CoV-2 viral load and reduction in clinical burden: the experience of a hospital in northern Italy. *Clin Microbiol Infect.* 2021;27(1):131.e1-3. <https://doi.org/10.1016/j.cmi.2020.10.006> PMID: 33059091
 11. Buchan BW, Hoff JS, Gmehlin CG, Perez A, Faron ML, Munoz-Price LS, et al. Distribution of SARS-CoV-2 PCR Cycle Threshold Values Provide Practical Insight Into Overall and Target-Specific Sensitivity Among Symptomatic Patients. *Am J Clin Pathol.* 2020;154(4):479-85. <https://doi.org/10.1093/ajcp/aqaa133> PMID: 32687186
 12. Binnicker MJ. Challenges and Controversies to Testing for COVID-19. *J Clin Microbiol.* 2020;58(11):e01695-20. <https://doi.org/10.1128/JCM.01695-20> PMID: 32817231

License, supplementary material and copyright

This is an open-access article distributed under the terms of the Creative Commons Attribution (CC BY 4.0) Licence. You may share and adapt the material, but must give appropriate credit to the source, provide a link to the licence and indicate if changes were made.

Any supplementary material referenced in the article can be found in the online version.

This article is copyright of the authors or their affiliated institutions, 2021.